

Declaration of Conformity

for the: **Flexi-therm**



Flexi-therm


The
directives:

laboratory heating device satisfies the following standards and

- BSEN 61010:2010 satisfying the Low Voltage Directive 2014/35/EU
- BSEN 61326:2013 satisfying the EMC directive 2014/30/EU
- BSEN 61010-2-010:2014
- UL 61010-1:2001 2nd Edition/CAN C22.2 CSA 61010-1
- FCC CFR 47 Parts 15.107 and 15.109, Class B
- METlabs NRTL Safety Approval – Listing No E113628

Approved for and on behalf of Integrated Technologies Ltd by:

Andrew Goodenough



Quality Assurance and Regulatory Affairs Manager

Date 12-JUNE-18